



StacSense

User manual

Version 1.3



For Sales & Service Contact

geotech

2650 E. 40th Ave. • Denver, CO 80205
Phone 303-320-4764 • Fax 303-322-7242

1-800-833-7958

www.geotechenv.com

The most recent version of this manual is available on our website:

<https://www.aqualabo.fr/>

IMPORTANT:

The information contained in this document may be patented or under patent application by Aqualabo. Possession of this document does not confer any license on these patents.

This manual is updated periodically. Updates are included in the new editions.

All information provided in this edition of the manual may be modified before the products described become available.

All rights reserved.

Copyright © 2019, Aqualabo

Reproduction in whole or in part is prohibited without Aqualabo's written consent.

CONTENTS

1. Warranty	6
2. Information	7
3. Safety	8
3.1 Safety instructions	8
3.1.1 Safety instructions in the manual	8
3.1.2 Labels on the product	8
3.2 OPERATING SAFETY	8
3.2.1 Intended use	8
3.2.2 Unauthorized use	8
3.2.3 User qualification	8
3.3 Notes on handling	9
3.4 Packaging	9
4. Overview OF THE STACSENSE SENSOR	10
4.1 Package contents	10
4.2 General view OF THE STACSENSE SENSOR	11
4.3 MEASURING PRINCIPLE	12
4.4	13
5. Connection and wiring	14
6. CONDITIONS OF USE	15
6.1 SITUATION SIMULATION	15
6.2 SENSOR FASTENING	15
6.3 LOCALIZED AIR INJECTION SYSTEM	16
6.4 UPKEEP OF THE SENSOR	17
7. COMMUNICATION	17
7.1 GENERAL POINTS	17
7.1.1 Communication and protocol	17
7.1.2 Sensor address:	18
7.2 SETTINGS	18
7.3 MEMORY PLAN: LIST OF MAIN REGISTERS	19
7.4 Calibration	20
7.4.1 Actions required to calibrate a parameter:	20
7.4.2 Fluid temperature setting:	20
7.4.3 Reference optical signal:	20
7.4.4 COD, BOD or TOC equivalence	21
8. ANNEX : Using the CALSENS SOFTWARE	23
8.1 CALSENS, version 1.4: general information	23
8.2 Equipment required	23
8.3 Installation	23

StacSense

8.3.1 Installation of the Calsens software tool	23
If you have an older version of Calsens	23
8.3.2 Installation of drivers	24
Missing DLL	26
8.4 SCAN function	27
8.4.1 Conflict of address:	28
8.4.2 Managing an undetected sensor:	28
8.5 Sensor detail	29
8.6 Calibration	30
8.6.1 Temperature	30
8.6.2 SAC254, reference signal in clear water	30
8.6.3 Coefficients for COD, BOD or TOC equivalence	33
8.7 Real-time view	34

1. WARRANTY

The STACSENSE probe equipment sold by the AQUALABO company are guaranteed against all manufacturing defects for a period of 2 year excluding consumables (unless expressly stipulated by AQUALABO) from:

- the date on which the buyer or authorized representative declares technical acceptance of the equipment at the factory,
- or alternatively,
 - for mainland France: from the date of the delivery note,
 - for other destinations: from the date of shipment certified by LTA, way bill, or bill of lading.

AQUALABO's warranty applies exclusively to flaws resulting from a design or inherent defect. It is strictly limited to the free delivery of replacement parts (except consumables) or the repair of the device in our workshops within 10 working days, transport not included.

The following are, by express agreement, formally excluded from our warranty:

- Any economic damage, such as staff costs, loss of profit, business disruption, etc.
- Any failure due to improper use of the device (unsuitable mains power, fall, attempted conversion, etc.), lack of maintenance by the user or poor storage conditions.
- Any failure due to using parts, not supplied by AQUALABO, on its equipment.
- Any failure due to transporting equipment in non-original packaging.
- Batteries, aerals, and in general, any item listed under "accessories".

Our customers are asked to always request our agreement before sending us a device for repair. No returns will be accepted without the prior written consent from our after-sales service which will stipulate the return procedure. In this case, the items will be returned in their original packaging, carriage paid, to the following address:

AQUALABO - 115 Rue Michel Marion 56850 Caudan - France

We reserve the right to reship any device received without this agreement. Regardless of the type and conditions of transport chosen to ship the equipment for repair under guarantee, in the original packaging, the related expenses as well as insurance costs will be the customer's responsibility.

Any damage resulting from returning the equipment falls within the framework of the guarantee on the express condition that customers have sent their claims to the carrier, by registered letter with acknowledgment of receipt, within forty-eight hours, with a copy of the letter sent to AQUALABO.

For devices with a warranty form, it only applies if the form received with the device is returned to AQUALABO duly completed.

SOFTWARE WARRANTY

The software is guaranteed by the software publisher or distributor under the conditions specified in the documentation related to those software packages.

AQUALABO does not, under any circumstances, guarantee the software packages.

By express agreement, we formally exclude from our warranty all economic damage, in particular for staff costs, loss of profit, business disruption, etc.

Customers are informed that AQUALABO may not, under any circumstances, be held responsible for any failure or bugs the software contain.

PROPERTY RIGHTS AND TRADE SECRETS

This document is protected by AQUALABO copyright (c) 2019 and all copyright is expressly reserved. The software supplied with the device, or referred to, contains trade secrets and confidential information which are AQUALABO property. It is legally protected by international copyright (c) law.

AQUALABO affords the user a license to use its software. This may not be divulged, used or duplicated for back-up purposes without written permission from AQUALABO. The beneficiary must attach a copy of this document to any fully or partially authorized reproduction.

2. INFORMATION

AQUALABO equipment has been designed, manufactured, tested and inspected in accordance with ISO 9001 procedures.

If the equipment is not used immediately, it should be stored in a clean, dry place. Abide by the following storage temperatures (10-35°C).

AQUALABO equipment is carefully inspected before packaging. Upon receipt of your device, check the condition of the packaging and if you notice an anomaly, submit the usual reservations with the carrier **within 48 hours**. Then consult the packing list and check that everything is in order. Finally, if you notice that something is missing or the equipment damaged, contact AQUALABO without delay.

STACSENSE sensors are entirely designed and manufactured by AQUALABO in France.



AQUALABO

90 Rue du Professeur Paul Milliez • 94506 Champigny sur Marne • France

Tel.: +33 (0)1 55 09 10 10

Hotline: +33(0)4 11 71 97 41

E-mail: info@aqualabo.fr

Website: <https://www.aqualabo.fr/>

3. SAFETY

3.1 SAFETY INSTRUCTIONS

3.1.1 SAFETY INSTRUCTIONS IN THE MANUAL

This manual provides important information on how to operate the product safely. Read it carefully to familiarize yourself with the product before commissioning and using it. This manual should be kept near the product so that the information you need is always easy to find.

Important safety instructions are highlighted in this manual. They are indicated by the warning symbol (yellow triangle) on the left-hand side.

CAUTION



Indicates a potentially hazardous situation that may result in serious (irreversible) injury, death, product damage, or destruction if the safety instructions are not followed.

3.1.2 LABELS ON THE PRODUCT

Be aware of all information labels and safety symbols affixed to the product.

3.2 OPERATING SAFETY

3.2.1 INTENDED USE

Comply with the following points for safe operation:

- Store and use the device under the environmental conditions mentioned in this manual ([see Characteristics](#))
- Do not disassemble the device
- Power the device in compliance with the specified DC voltage range.
- Comply with the use restrictions given below

3.2.2 UNAUTHORIZED USE

The product must not be operated if:

- it is visibly damaged (e.g. after being transported)
- it has been stored under adverse conditions for a long period of time

3.2.3 USER QUALIFICATION

We assume that operators know how to manage this equipment by virtue of their professional training and experience. In particular, operators must be able to understand and correctly implement the safety instructions/labels when using the product. Trained personnel should be familiar with and follow the instructions in this manual.

3.3 NOTES ON HANDLING

StacSense sensors are electronic devices. As such, they must be treated with care. Always protect the device from conditions that could adversely affect its components. In particular, adhere to the following points:

- The ambient temperature and humidity, when using and storing, must be within the limits indicated in the TECHNICAL CHARACTERISTICS section.
- Regardless of the circumstances, the device must be protected from the following influences:
 - Intensive exposure to light and heat
 - Corrosive or caustic fluids, organic solvents or vapor with a high solvent content.
- This sensor is an optical measurement device. As a result, the optical path and more particularly the two optical windows must be protected. Dirt and scratches may affect measurement performance.
- All work on the inside of the instrument, must be performed by AQUALABO or by AQUALABO-approved technicians.

3.4 PACKAGING

The StacSense sensor is shipped in packaging designed to protect it during transport. It is essential that you keep the original packaging as well as the inner packing in such a way as to ensure optimum protection of the device from impact in the event of transport. The original packaging is also required for return transport under appropriate conditions if repair is needed. Please remember that we shall decline any warranty claim for damage caused by unsuitable transportation.

4. OVERVIEW OF THE STACSENSE SENSOR

4.1 PACKAGE CONTENTS

You have just received your sensor:



Model shown: 2 mm optical path

The package consists of a StacSense sensor with a single power and communication cable. A USB stick containing this manual and all the product-related information.
Cable length varies depending on the model ordered.

Upon receipt of your device, check the condition of the packaging and if you notice an anomaly, submit the usual reservations with the carrier **within 48 hours**. If you notice that something is missing or the equipment damaged, contact AQUALABO without delay.



CAUTION:

The sensor and cable (along its entire length) must be in perfect condition before immersion in the fluid studied.

4.2 GENERAL VIEW OF THE STACSENSE SENSOR

The STACSENSE sensor has a largely cylindrical shape extended by a fastening ring or grip, which protects the cable outlet area.

The optical path is determined at the time of order, depending on the sensor's field of use.

The optical windows are located on two separate mechanical components.

A female thread located on the opposite side of the main measuring area adapts to a compressed air supply connection.



General views - 2mm optical path sensor

General views - 50mm optical path sensor



CAUTION:

The cable is used to power the sensor and communicate with the master device. The product's seal is not guaranteed when the sensor is suspended by its cable or when the cable is damaged (cut or partial abrasion of the outer sheath).



CAUTION:

The STACSENSE sensor outer shell is made of 316L stainless steel. This material is not compatible with use in salt water (seawater, spray): risk of corrosion in the short term.

4.3 MEASURING PRINCIPLE

The STACSENSE sensor is an optical sensor for measuring absorption in aqueous solution (natural or wastewater). The optical system consists of five major components:

- Main optical source, UV-C LED, peak wavelength 255 +/- 5nm,
- Secondary optical source, visible LED, peak wavelength 530 +/- 5nm,
- Assembly of optical components to manage light beams
- Optical path through the observed fluid,
- Detector, photo-diode.

The light emitted by the source is guided through the optical windows and fluid studied. The latter absorbs part of the radiation. The intensity of the light I captured by the optical receiver is used. The reference light intensity I_0 is obtained in a sample of clear, pure water at a stable temperature.

Transmittance T is defined by the following equation: $T = I/I_0$

Absorbance A is defined by the equation: $A = -\log_{10} T$

The SAC 254 parameter is obtained by using absorbance at 254 nm as well as at 530 nm and the optical path length L expressed in meters.

$$\text{SAC 254} = [A_{(254\text{nm})} - AT_{(530\text{nm})}] / L$$

NOTE: reference document DIN 38404-3; July 2005

NOTE: Absorbance at 530nm is used to perform a turbidity correction. The initial light beam is mitigated by the scattering of light particles and the insoluble substances in the medium analyzed.

CAUTION:



The STACSENSE sensor has an optical system based on a UV-C light source. Never look directly into the light source. Risk of irreversible damage to sight.

CAUTION:



An external, intense light source may saturate the sensor's optical system receiver and therefore distort the measurement.

4.4 CHARACTERISTICS

Digital communications	RS485, MODBUS RTU protocol The STACSENSE sensor is a modbus slave. The master device configures and commands the measurement then records the result(s) by reading logs defined in the magnetic core plane.	SDI-12 Measurements made by the sensor are also accessible via the SDI-12 protocol.	
Class	Overall dimensions Diameter: 48mm Length: <i>see footprint diagram</i> . 2mm optical path: 50mm optical path:	Weight 1600 - 1800 g according to the optical path (cable not included) Materials <ul style="list-style-type: none"> • 316 L stainless steel • Corning 7980 • O-rings: Fluoro-elastomer (FPM / FKM) 	Connections Standard cable length: 3, 7 and 15 meters. Bare wires or 6-male-contact plugs.
Power Supply	5.4 V ^{1,2} at 26 V ³ DC ¹ Absolute minimum 5.2 V with 1 m of cable ² Minimum voltage subject to cable length-related losses ³ 28.0 V absolute maximum		
Environment	Ingress protection rating IP68	Temperature Storage: -10°C to + 50°C Use: 0°C to +40°C	Standards Compliant to EC marking
Identification	The product serial number is written on the outer casing of the sensor.	The sensor is identified by a label affixed to the outer casing.	
Accessible settings	Standard settings Fluid temperature (°C) Spectral Absorption Coefficient, SAC 254 (m-1) COD equivalent (mg/L)* BOD equivalent (mg/L)* TOC equivalent (mg/L)* * The conversion coefficients are determined in the factory from aqueous solutions of potassium hydrogen phthalate (KHP) under laboratory conditions.	Extended settings Turbidity (530nm) (FAU) ** Transmittance and Absorbance. ** The equivalent turbidity is determined in the factory, from aqueous suspension of formazine, under laboratory conditions.	

5. CONNECTION AND WIRING

The STACSENSE sensor comes with a factory-installed cable and connector.



CAUTION:


The modification of the connector installed by the manufacturer, bare wire or 6-contact plug, represents a major transformation of the product and entails a loss of guarantee.

Never exceed a voltage of 10VDC (absolute maximum rating) on communication lines RS485, A or B, under penalty of irreversible destruction of the transceiver component RS 485.

SDI-12: respect the voltage value described in the associated standard (nominal: 5 VDC)

Always connect ground + shield first.

Conductor IDENTIFICATION TABLE:




Cable length up to 15m included		Cable length from 15 to 100 m	
1 - Red	Power Supply, V+	Red	Power Supply, V+
2 - Blue	SDI-12	Yellow	
3 - Black	Weight	Orange	
4 - Green	B "RS-485"	Purple	
5 - White	A "RS-485"	Pink	
6 - Green/Yellow	Cable shielding	2 - Blue	SDI-12
Connect wires 3 and 6 together		3 - Black	Weight
		4 - Green	B "RS-485"
		5 - White	A "RS-485"
		6 - Green/Yellow	Cable shielding
		Connect wires 3 and 6 together	

6. CONDITIONS OF USE

6.1 SITUATION SIMULATION

The STACSENSE sensor may be fully immersed. In the case of partial immersion, the entire optical path must be kept in the fluid studied.

Lateral flow is recommended to prevent particles from building up, deposits or gas bubbles in the measuring zone between the optical windows.



6.2 SENSOR FASTENING

A sturdy system is required to hold the sensor, in order to:

- keep the sensor in the circulating fluid, according to the appropriate position, away from any sedimentation zone and thus obtain representative measurements of the medium,
- avoid impacts on the sensor, contact with the walls, bottom of the tank, floor or any other object in the fluid.
- eliminate mechanical strain on the power and communication cable.

CAUTION:




The power and communication cable shall not be used to move or transport the sensor. The ring located at the end of the sensor is designed for this purpose.

As an example, a clamp consisting of two half-shells with a stainless steel base (see illustration below, 48.3mm diameter model) can be used to permanently position the sensor in the fluid analyzed. This mechanical system using a polypropylene shell, completely isolates the body of the sensor from its support.




*Mechanical sensor-holding system
(front view)*




*Clamp (side view) - D48.3mm
compatible with the sensor*

6.3 LOCALIZED AIR INJECTION SYSTEM

A female thread (G1/8) located on the opposite side of the main measuring zone is used to connect a compressed air supply. This air injection device, located near the two optical windows removes residue from the measurement zone.



Overview view, rear of the sensor



Detailed view of the compressed air circuit

NOTE:

Using a compressed air supply quick-release connector, bent to 90°, compatible with a standard semi-flexible hose, and with a 6mm external diameter is recommended. The 90° elbow connects the air supply tube along the sensor body to the communication cable.



50mm optical path sensor with 90° elbow connection



CAUTION:

Installing a compressed air quick-release coupling, made of metal other than stainless steel, may generate corrosion phenomena locally that damage the sensor.

6.4 UPKEEP OF THE SENSOR

The STACSENSE sensor is an optical sensor which is sensitive to pollution in the measuring zone. The user must, during maintenance operations, check the status of the two optical windows and identify any scratches that may require calibration in clear water. Routine servicing of the optical windows can be performed by rinsing in clear water, followed by cleaning with a clean rag soaked in isopropyl alcohol (IPA).



CAUTION:

The use of an abrasive cleaning product may permanently damage the optical windows.

The use of an organic solvent such as acetone may damage the product's seals as well as the cable sheath.

7. COMMUNICATION

7.1 GENERAL POINTS

7.1.1 COMMUNICATION AND PROTOCOL

The STACSENSE sensor has two digital serial communication modes:

- The Modbus RTU protocol (RS485 physical medium) enables the exchange of information between the master device and the sensor (slave) such as, the measurement configuration, measurement values and calibration of the available parameters.
- The SDI-12 protocol (TTL/5V Half-Duplex Series physical medium) triggers measurements and collects results from the main measurements.

7.1.2 SENSOR ADDRESS:

The StacSense sensor's RS485/Modbus communication address is factory set. Its value is constant (70) and different from that attributed to any other model in the DIGISENS range, for example, OPTOD sensor (10). The broadcast address (0) is aimed at a sensor whose address is unknown to change its Modbus address.

Similarly, the ex-factory default SDI-12 address is 7.

These addresses can be modified by an operator wishing to manage a network of several sensors, thereby avoiding any conflict.

7.2 SETTINGS

The parameters accessible via the STACSENSE sensor are fall into three categories:

I_Main parameters:

- a. Temperature of the fluid, expressed in °C,
- b. Spectral Absorption Coefficient SAC 254, based on absorbance at 254 and 530nm, expressed in m^{-1}

II_Parameters deduced according to a first degree polynomial law (offset and slope)

- a. Chemical Oxygen Demand (COD) equivalent, expressed in mg/L
- b. Biological Oxygen Demand (BOD) equivalent, expressed in mg/L
- c. Total Organic Carbon (TOC) equivalent, expressed in mg/L of carbon

III_Estimated parameter

Turbidity of the fluid based on absorbance at 530 nm of a formazine aqueous suspension.

NOTE: The conversion coefficients, offset and slope, for each of the three deduced parameters (COD, BOD, and TOC), are factory set by default. However, these coefficients are dependent on the application and type of fluid observed. They must be defined by the user during any lasting measurement campaign.

7.3 MEMORY PLAN: LIST OF MAIN REGISTERS

The document describing the memory layout, common to sensors in the DIGISENS range, is the reference for exchanges between the master device and the sensor(s) connected to it.

The StacSense sensor provides most of the communication aspects of DIGISENS sensors.

The key components for collecting measurements are the following:

ACTIONS / DATA	Modbus address (as a hexadecimal)	Read, write	VALUE	TYPE	UNIT
Order of measurement	0x0001	Write	FF	full	/
Time required for the sensor to refresh the measurements*.	0x00A4	Read	Response Time	full	milliseconds
fluid temperature	0x0053	Read	data	floating	° Celsius
parameter 1 SAC	0x0055	Read	data	floating	m ⁻¹
parameter 2 COD eq.	0x0057	Read	data	floating	mg/L
parameter 3 BOD eq.	0x0059	Read	data	floating	mg/L
parameter 4 COT eq.	0x005B	Read	data	floating	mg/L carbon

* The sensor is silent for this period of time, at the most, subsequent to the measurement request.

Extended measurement data is accessible over a second address range:

ACTIONS / DATA	Modbus address (as a hexadecimal)	Read, write	VALUE	TYPE	UNIT
fluid temperature	0x007F	Read	data	floating	° Celsius
parameter 1 SAC	0x0081	Read	data	floating	m ⁻¹
parameter 2 DCO eq.	0x0083	Read	data	floating	mg/L
parameter 3 DBO eq.	0x0085	Read	data	floating	mg/L
parameter 4 COT eq.	0x0087	Read	data	floating	mg/L carbon
UV Compensated absorbance.	0x0089	Read	data	floating	<i>no units</i>
GR Compensated absorbance	0x008B	Read	data	floating	<i>no units</i>
Turbidity eq.	0x008D	Read	data	floating	FAU
UV transmittance	0x008F	Read	data	floating	<i>no units</i>
GR transmittance	0x0091	Read	data	floating	<i>no units</i>

Warning: No Modbus communication on the BUS during measurement (between the measurement request and the end of the time required for measurement) and to no devices connected to the BUS

7.4 CALIBRATION

7.4.1 ACTIONS REQUIRED TO CALIBRATE A PARAMETER:

General case in 5 steps:

1. Restore the factory coefficients (optional),
2. Make measurements until a sufficiently stable value is obtained,
3. Write the known standard value related to this step,
4. If the calibration requires several steps, repeat actions 2 and 3 for the other coefficients,
5. Validate the coefficients of the parameter by writing the date and operator's name.

7.4.2 FLUID TEMPERATURE SETTING:

7.4.2.1 General points:

Sequence in two steps, i.e. two independent conditions for stabilized temperature. The operator must have a reference thermometer.

7.4.2.2 Detailed calibration of the temperature setting:

Temperature Calibration					
Step	Modbus Name	Calibrated Parameter	Description	Sensor conditions	Reference Input Value
1	Temp. Offset Coeff.	Temperature	Offset (°C) (CTN)	Water T1	T1 °C
2	Temp. Slope Coeff.	Temperature	Slope (%) (CTN)	Water T2	T2 °C

Step 1:

Place the sensor under steady low temperature conditions, fluid at T1. The master sends requests for repeated measurements until a sufficiently stable measurement is obtained, Write the temperature value T1 (°C), delivered by the reference equipment, in the "Standard" Modbus register of the sensor, related to the temperature offset. The temporary offset coefficient generated by the sensor is activated for the next step.

Step 2:

Place the sensor under stabilized upper temperature conditions, fluid at T2, significantly different from T1. The master sends requests for repeated measurements until a sufficiently stable measurement is obtained, Write the temperature value T2 (°C), delivered by the reference equipment, in the "standard" Modbus register of the sensor, related to the temperature slope. The slope coefficient is generated by the sensor.

Validation:

The coefficients generated by the sensor are considered to be temporary until final validation. This action is performed by the master by writing the operator's name and the date. The temporary coefficients will also have to be cleaned using the appropriate Modbus commands.

7.4.3 REFERENCE OPTICAL SIGNAL:

7.4.3.1 General points

This is a measurement in clear, pollution-free water, without bubbles or particles in suspension, at stabilized temperature.

7.4.3.2 Detailed calibration of the SAC 254 main parameter:

Clear water: SAC254 Calibration (UV and GR Transmittance Calibration)					
Step	Modbus Name	Calibrated Parameter	Description	Sensor conditions	Reference Input Value
1	Coeff 1	SAC 254 (Param. 1)	UV User Clear Water Tx Comp Slope	Clear water	0
2	Coeff 3	SAC254 (Param.1)	GR User Clear Water Tx Comp Slope	Clear water	0

Step 1: Zero adjustment of the UV light source (UV absorbance at 0, UV transmittance at 1)

Place the sensor under the following conditions: clear water stabilized at temperature T, degassed.

StacSense

The master sends requests for repeated measurements until a sufficiently stable measurement is obtained, Write the value Zero (\Leftrightarrow clear water) as the standard value of coefficient 1, in the related Modbus register. The coefficient required to adjust the UV zero of the sensor is calculated automatically by the sensor.

Step 2: Zero adjustment of the green light source (Green absorbance at 0, Green transmittance at 1) Immediately after step 1, write the value Zero (\Leftrightarrow clear water) as the standard value of coefficient 3, in the related Modbus register. The coefficient needed to set the sensor's Green zero is calculated automatically by the sensor.

Validation:

The coefficients generated by the sensor are considered to be temporary until final validation. This action is performed by the master by writing the operator's name and date of parameter 1. The temporary coefficients will also have to be cleaned using the appropriate Modbus commands.

7.4.4 COD, BOD OR TOC EQUIVALENCE

7.4.4.1 General points

SAC 254 is a parameter for organic substances dissolved in water that absorb UV radiation. It provides information on water contamination. Despite the similarities, the parameters can not always be inter-converted. However, a correlation can be established between the SAC 254 parameter and another parameter such as TOC or COD. The STACSENSE sensor then provides equivalence data.

To obtain the correlation, it is recommended to measure the SAC for a few days on samples of polluted water which will also be analyzed using laboratory measuring equipment according to the standardized method. The conditions for obtaining useful data for efficient conversion represent daily monitoring with periods of low and high loads, as in the case of urban effluent. During these peak periods, you have to:

- read the SAC value delivered by the StacSense sensor,
- take a representative fluid sample at the sensor location,
- stabilize and store successive samples at 4°C until analysis,
- perform laboratory analyses of the parameter to be correlated.
- use the data to determine a conversion law.

COD, BOD or TOC equivalences are calculated directly in the sensor according to a first degree law. This pair of coefficients (offset and slope) is specific to each parameter.

Example: COD eq. = **Slope_{COD}** * SAC254 + **Offset_{COD}**

The default conversion coefficients, filled in at the factory, are as follows:

- DCOeq = 1,81 * SAC254 + 0,0
- DBOeq = 0,48 * SAC254 + 0,0
- TOCeq = 0,69 * SAC254 + 0,0

As per the usual calibration process, the user value of the coefficient, offset or slope, such as a reference input value, is entered in a sensor register instead of the factory coefficient.

The user evolves this linear relationship according to the results of a significant measurement campaign with two techniques in parallel, the UV sensor and the laboratory analysis of samples taken regularly, in the same place.

Example of experimental determination of the conversion coefficient SAC to COD (50mm sensor optical path):

At the beginning of the measurement and sampling campaign, the conversion coefficient SAC254 to factory COD is active (slope value at 1.81; offset zero).

Average COD value provided by the sensor = 36mg/L
Mean laboratory sample test result = 22 mg/L

The conversion slope SAC254 to COD is therefore adjusted in the sensor such as $1,81 \times 22 / 36 = 1,11$

As an example, this coefficient equal to 1.11 was determined on water leaving the treatment plant.

7.4.4.2 COD Equivalence

COD customer-user coefficient input					
Step	Modbus Name	Calibrated Parameter	Description	Sensor conditions	Reference Input Value
1	Coeff 11	COD (Param. 2)	COD Calculation Offset	Not required	Offset Value
2	Coeff 12	COD (Param. 2)	COD Calculation Slope	Not required	Slope Value

Step 1: No special operating condition. The master device enters the coefficients that are useful for the law of conversion.

Write the offset value as a Reference input value related to coefficient 11.
The sensor automatically copies the offset value written in the standard into the related coefficient memory box.

Step 2: Immediately after step 1, write the slope value as a Reference input value related to coefficient 12.
The sensor automatically copies the slope value written in the standard into the related coefficient memory box.

Validation:

Write the operator name and date of parameter 2 to validate the user coefficients needed to obtain the COD equivalence. The temporary coefficients will also have to be cleaned using the appropriate Modbus commands.

7.4.4.3 BOD Equivalence

BOD customer-user coefficient input					
Step	Modbus Name	Calibrated Parameter	Description	Sensor conditions	Reference Input Value
1	Coeff 13	BOD (Param. 3)	BOD Calculation Offset	Not required	Offset Value
2	Coeff 14	BOD (Param. 3)	BOD Calculation Slope	Not required	Slope Value

Step 1: No special operating condition. The master device enters the coefficients that are useful for the law of conversion.
Write the offset value as a Reference input value related to coefficient 13.
The sensor automatically copies the offset value written in the standard into the related coefficient memory box.

Step 2: Immediately after step 1, write the slope value as a Reference input value related to coefficient 14.
The sensor automatically copies the slope value written in the standard into the related coefficient memory box.

Validation:

Write the operator name and date of parameter 3 to validate the user coefficients needed to obtain the BOD equivalence. The temporary coefficients will also have to be cleaned using the appropriate Modbus commands.

7.4.4.4 TOC equivalence

TOC customer-user coefficient input					
Step	Modbus Name	Calibrated Parameter	Description	Sensor conditions	Reference Input Value
1	Coeff 15	TOC (Param. 4)	TOC Calculation Offset	Not required	Offset Value
2	Coeff 16	TOC (Param. 4)	TOC Calculation Slope	Not required	Slope Value

Step 1: No special operating conditions. The master device enters the coefficients that are useful for the law of conversion.
Write the offset value as a Reference input value related to coefficient 15.
The sensor automatically copies the offset value written in the standard into the related coefficient memory box.

Step 2: Immediately after step 1, write the slope value as a Reference input value related to coefficient 16.
The sensor automatically copies the slope value written in the standard into the related coefficient memory box.

Validation:

Write the operator name and date of parameter 4 to validate the user coefficients needed to obtain the TOC equivalence. The temporary coefficients will also have to be cleaned using the appropriate Modbus commands.

8. ANNEX : USING THE CALSENS SOFTWARE

8.1 CALSENS, VERSION 1.4: GENERAL INFORMATION

The CALSENS software tool manages all the sensors in the DIGISENS range.
Version 1.4 also supports the StacSense UV probe.

This version specifically manages:

- a) Extended parameter support for sensors that provide more than 5 parameters,
- b) The system of editable parameters to directly enter an offset and a slope for specific StacSense probe parameters,
- c) The real-time view supports a larger number of parameters including UV and visible green (530nm) absorbance and transmittance.



/ Click this icon to see details about the sensor and settings.

8.2 EQUIPMENT REQUIRED

- Windows PC
- RS485/USB converter, reference model: RS485/USB converter NC-FIX-C-00020, 1 MODULE 4201 sensor input
- 12VDCA power supply

8.3 INSTALLATION

The installation involves 2 steps:

- 1 software installation;
- 2 driver installation for the FTDI USB adapter cable.


8.3.1 INSTALLATION OF THE CALSENS SOFTWARE TOOL

To install Calsens, run the “**Calsens Installer.msi**” and follow the instructions.

Calsens Installer.msi	29/10/2019 16:00	Package Windows...	6 835 Ko
setup.exe	29/10/2019 16:00	Application	444 Ko

IF YOU HAVE AN OLDER VERSION OF CALSENS


You need to uninstall the older version. Go to the “**Control Panel**” in Windows. Click on “**Uninstall a program**” in “**Program**”.




Then search for “**Calsens**” in the search bar. Then do a right click on “**Calsens**” > “**Uninstall**” and follow the uninstall instructions. Finally install Calsens as previously stated.

8.3.2 INSTALLATION OF DRIVERS

To install the drivers, open the “**peripheral device manager**” and select “**PONSEL MODBUS Adapter**”




Right click on “**PONSEL MODBUS Adapter**” and click “**update the driver**”.




Click "Find a driver on my computer".

Select the "PONSEL USB Driver WHQL 2.12.28" folder on the installation CD to install the correct drivers.



Once the driver is installed, Calsens is ready to use.



MISSING DLL


When Calsens is started and you try to use USB port for Modbus for the first time, the software may indicate that the DLL **ftd2xx.dll** is missing. It's needed to copy the right DLL in the Calsens repertory. To do this action you have to determine your system's architecture : 32 bits or 64 bits (Computer > right click > properties > system type).

- If the system is 64 bits, you have to copy the DLL **ftd2xx64.dll** present in folder **PONSEL USB Drivers WHQL 2.08.28\amd64** to the Calsens Application folder. Then you need to rename it in **ftd2xx.dll**.
- If the system is 32 bits, you need to copy the DLL **ftd2xx64.dll** present in folder **PONSEL USB Drivers WHQL 2.08.28\i386** to the Calsens Application folder.


When these actions are done, you can re start Calsens and use normally the PONSEL USB Cable for Modbus.

8.4 SCAN FUNCTION

To detect the StacSense probe, scan the Modbus addresses using the SCANNER button on the CALSENS interface. This operation consists in testing each RS485/Modbus network address in order to detect connected, compatible sensors.



8.4.1 CONFLICT OF ADDRESS:



/

When two sensors have the same address there is a conflict. The conflict can be resolved automatically or manually. The automatic method requires the sensors to be connected one by one so that CALSENS can re-address the sensors. The manual method specifies the serial number of each sensor.


8.4.2 MANAGING AN UNDETECTED SENSOR:

If the StacSense probe is not detected, several things have to be checked:

1. Connection status of the probe cable with the USB/RS485 converter,
2. Sensor 12VDC power supply connection status,
3. If the power supply is enabled,
4. If the RS485 A/B communication lines are used properly.

8.5 SENSOR DETAIL

The sensor detail shows the StacSense probe's parameters. The magnifying glass icon  / shows the details of the parameters and indicates which parameters can be calibrated.



SN-PUVTA-0014 sensor detail

Serial Number : SN-PUVTA-0014

Description : UVT254/Temperature PONSEL

Firmware version : 0.05 

Hardware version : 0.01

Modbus address : 64 

SDI12 address : 7 


Parameter	Value	Unit	Status	Range	Detail
Temperature	24.90	°C		0/40 °C	
SAC 254	0.57	1/m		0/X 1/m	
COD equivalent	1.03	mg/l			
BOD equivalent	0.27	mg/l			
TOC equivalent	0.41	mg/l			

Back Refresh measures

8.6 CALIBRATION

Two parameters can be calibrated on the UV probe: temperature and SAC 254 using the "Calibrate" button on the main screen. Additional parameters: COD, BOD, or TOC equivalence can be configured via the parameters "detail" screen.

Selecting the parameter to be calibrated:



8.6.1 TEMPERATURE

Similar to other sensors in the DIGISENS range, see the documentation built into the CALSENS software tool.

8.6.2 SAC254, REFERENCE SIGNAL IN CLEAR WATER


The calibration of SAC 254 is useful when adjusting the optical reference signal. Place the sensor in clear, temperature-stabilized water. Eliminate any bubbles trapped in the measurement zone (optical path of the sensor). Follow steps 1 and 2. Wait a few seconds at each step until the values stabilize.

Step 1:

In this step, the UV probe emits UV radiation at a wavelength of 254 nm to calibrate the zero absorbance.



CAUTION:
UV radiation may cause irreversible damage to eyesight.




Note:

As shown in the above figure, the CALSENS tool manages single or multiple calibration. In the second case, a set of identical sensors can be adjusted simultaneously in the same solution.

Step 2:

In this step, the UV probe emits visible radiation at a wavelength of 530 nm to calibrate the zero absorbance. The measurement status is visible as well as the stability.



/

Assessment:

The assessment step validates the temporary coefficients defined during the previous two steps. Check that the coefficients are consistent before validating them.


After validating the calibration (VALIDATE CALIBRATION button), SAC 254 is calibrated.

8.6.3 COEFFICIENTS FOR COD, BOD OR TOC EQUIVALENCE

There is no calibration step for COD, BOD and TOC coefficients. You can enter the offset and slope values directly in the parameter "detail" view.


///

This component (see area outlined in red) replaces the list of calibration logs. In the example below, set the desired values and validate to write the appropriate values of the fine-tuned conversion coefficients, in the sensor, from SAC254 to COD.



8.7 REAL-TIME VIEW

The 5 main parameters (Temperature, SAC 254, COD, BOD or TOC) can be seen in the real-time view as well as the extended parameters (turbidity, UV absorbance and UV transmittance).



Select the desired parameters using the boxes in the "Display" tab. The display window is then made lighter to display data from several sensors.

/
When the desired parameters are selected, values can be stored in a CSV file, named at the beginning of the acquisition campaign (START THE LOG button).

Note:
Only the selected settings will be saved.

CALL GEOTECH TODAY (800) 833-7958

Geotech Environmental Equipment, Inc.
2650 East 40th Avenue • Denver, Colorado 80205
(303) 320-4764 • FAX (303) 322-7242
email: sales@geotechenv.com • website: www.geotechenv.com

AQUALABO
eur Paul Milliez • 94506 Champigny sur Marne • France
Tel.: +33 (0)1 55 09 10 10
Hotline: +33(0)4 11 71 97 41
E-mail: info@aqualabo.fr
Website: <https://www.aqualabo.fr/>